

Centrifuge

(For GCRI, Ahmedbad)

TECHNICAL BID

Name of Tenderer: _____

Technical Specifications		Yes / No
Conformity to standards	CE	
Certificate number and date	IN34247A Dt. 29.01.2019	
Performance Parameters		
Purpose	For sedimentation of body fluids	
Centrifuge Design	Bench Top	
Timer provided	Yes	
Maximum RPM	2000- 3000	
Type of Motor	Brushless (without carbons)	
Facility to ensure that In case Machine is running and lid is opened with manual lock by user , machine shall stop	Yes	
Timer Set range	60 minutes	
Display	LED or LCD	
Speed Controller	Microprocessor controlled pre programmed speed controller	
Rotor configuration	Swing Out Rotor and changeable rotor	
Maximum Volume of the tube testable	50 ml centrifuges tubes	
Rotor Size	At a time 50 x 8 ml and 12x15 ml	
Power supply	Single phase, 230 V 50 HZ AC Source	
Power consumption in watts	300	
The Equipment shall be installed and demonstrated before Acceptance installation and demonstration	demonstration required	
Warranty in Years	1	
Reports		
ISO certification for Manufacturer	Yes	
Certification number and date , if no certificate provided put NA	ISO 13485 Certificate no 910618	
Copies of all certifications and reports to be provided to buyer on demand at time of supplies	Yes	

- we need rotor size in which we can centrifuge 2 size tubes at a time (i.e.- 15 ml. And 50ml.)

Remark (If any): _____

Important Note :

- 1. Party is requested to submit their technical and commercial bid in separate envelopes on or before Dt. 30/03/2021 @ 17:00pm.**
- 2. The tenderer is requested to submit catalogue, literature, brochure of quoted model.**

Sign and Stamp of Tenderer

Centrifuge

Commercial Bid

Name of Tenderer: _____

Sr. No.	Particular	Amount (Rs.)
A	Basic Price of Equipment	
B	Applicable Tax (eg.GST etc.)	
C=A+B	Total Equipment's Cost	

* For L1 calculation only basic price.

* Tax paid at actual at the time of invoice.

Sign and Stamp of Tenderer