CORRIGENDUM-2

Tender Id:536234

The Placement of Fully Automated Real Time PCR for Microbiology on Reagent Rental for Five Years

| Uploaded Specifications | Read As |
|---|--|
| System for Real-Time Amplification | |
| Point No:13 | Point No:13 |
| The capacity of system should be 96 wells | The capacity of system up to 96 wells. |
| or more | |
| Point No:16 | Point No:16 |
| | |
| Adjustable hot lid temperature. | Adjustable hot lid temperature. (Optional) |
| | |
| | |
| Tender Closing Date: | Tender Closing Date: |
| 28-07-2022 | 04-08-2022 |
| | |
| | |
| | |